


3-1 Rotational/Circumferential Speed: f1



FUNCTION

Application example



Basic Operation

Multiplies the input frequency (Hz) of INA by 60 and displays the result in rpm. When the appropriate prescale value is selected, the rotational speed of the object is displayed. Obtain display value D as follows:

$$D = f_A \times 60 \times \alpha$$

f_A : Input frequency of INA (Hz)

α : Prescale value

INB input will be ignored.



SETTING

Item	Unit of display	Prescale value
Rotations	rpm	1/N
	rps	1/60N
Frequency of input pulse	Hz	1/60
	kHz	1/60000
Rotational speed	mm/s	$1000\pi d/60N$
	cm/s	$100\pi d/60N$
	m/s	$\pi d/60N$
	m/min	$\pi d/N$
	km/h	$0.06\pi d/N$

Where,

N: Number of pulses per rotation

πD : Length (m) of one rotation

Example:

Displaying rotations (rpm) on condition that there are two pulses per revolution.

$$\text{Prescaling value } (\alpha) = 1/2 = 0.5 = 5.0000 \times 10^{-1}$$

$$\text{Prescale value} = X \times 10^Y \text{ (X: mantissa, Y: exponent)}$$

X (mantissa) of input A = 5.0000
Y (exponent) of input A = -1



REFERENCE

Refer to 4-2 *Setting Mode*.

Hold Measured Value

When the HOLD input is turned ON, measurement stops and the input measured just before the HOLD input turned ON is held. While the HOLD input is ON, the K3NR holds display output, comparative output, and BCD output.

When the comparative output from the Output Board is connected to the HOLD input terminal, the value measured immediately after the occurrence of an error can be obtained.

Performance Characteristics

Accuracy of measurement	$\pm 0.006\% \text{ rdg} \pm 1 \text{ digit}$ (ambient temperature: $23^{\circ}\text{C} \pm 5^{\circ}\text{C}$)
Measurement range	Sensor with transistor output: 0.5 mHz to 50 kHz Sensor with relay output: 0.5 mHz to 30 Hz
ON/OFF pulse width	Sensor with transistor output: 9 μs min. Sensor with relay output: 15 ms min.

Response time	Output configuration				
	Relay output	Transistor output	BCD and transistor output	Linear and transistor output	Communication and transistor output
Comparative output	200 ms max.				
BCD output	---	---	Refer to page 139.	---	---
Linear output	---	---	---	220 ms max.	---

Available Functions

Available functions in this mode are indicated as “Yes” in the following table.

Menu	Function	Displayed Character	Availability	Reference page
---	Max./Min. value display and reset	---	Yes	121
	Estimated frequency calculation	---	Yes	153
sUset (See note 2)	Set value bank no. of set values	s.bank	Yes	60
	HH set value	sU*.hh		
	H set value	sU*. h		
	L set value	sU*. l		
	LL set value	sU*.ll		
pscl	Select bank no. of prescale value	p.bank	Yes	65
	Prescaling value of input A X (mantissa) Y (exponent)	ps*.ax ps*.ay	Yes	
	Prescaling value of input B X (mantissa) Y (exponent)	ps*.bx ps*.by	No	
	Decimal point position	dec.p.*	Yes	
setup	Operating mode	func	Yes	70
	Input A sensor type	ina	Yes	72
	Input B sensor type	inb	No	
	Auto zero time of input A X (mantissa) Y (exponent)	=ro.ax =ro.ay	Yes	74
	Auto zero time of input B X (mantissa) Y (exponent)	=ro.bx =ro.by	No	
	Display time unit	time	No	77
	Communications unit no. (See note 1)	u-no	Yes	79
	Baud rate (See note 1)	bps	Yes	
	Word length (See note 1)	len	Yes	82
	Stop bits (See note 1)	sbit	Yes	
opt	Parity bits (See note 1)	prty	Yes	85
	Process time for averaging measured value	aUg	Yes	
	Startup compensation time	stime	Yes	
	Power failure memory	memo	No	
	Hysteresis (See note 1)	hys	Yes	
	Comparative output pattern (See note 1)	c-out	Yes	
	H linear output range (See note 1)	lset.h	Yes	
	L Linear output range (See note 1)	lset.l	Yes	
	Remote/Local programming (See note 1)	r-l	Yes	101

- Note**
1. The availability of the parameters depends on the type of selected Output Board.
 2. The selected bank number will be displayed where an asterisk (*) appears.